Investigation of the neutron-gamma ray discrimination performance at low neutron energy of a solution-grown stilbene scintillator

Augusto Di Chicco^{1,3,*}, Michael Petit², Robert Jacqmin¹, Vincent Gressier², Brian Stout³

¹CEA Cadarache, DEN/DER, 13108 Saint-Paul-lez-Durance, France.

²IRSN Cadarache, SDOS/LMDN, 13108 Saint-Paul-lez-Durance, France.

³Aix-Marseille Université, Institut Fresnel – UMR 7249, 13397 Marseille, France.

Abstract — The results of experiments performed with a Ø25 x 25 mm solution-grown stilbene crystal in mono-energetic neutron fields in the 80-to-230 keV energy range are presented. The goal of the measurements, performed at the AMANDE facility, was to explore the capabilities of this organic scintillator to measure neutrons at the lowest possible energy with good pulse shape discrimination (PSD). The time of flight (TOF) technique was used in order to help with the neutron-gamma discrimination. The data are collected via a programmable digital acquisition (DAQ) system CAEN DT7530 using the software CoMPASS with the charge comparison method (CCM). The data are analysed using post-processing codes developed in the ROOT environment. The results show that the stilbene detector has discrimination capabilities for energies as low as 80 keV.

Keywords — PSD, Organic scintillators, Solution-grown stilbene, Mono-energetic neutrons, Time of flight, digital acquisition system, AMANDE facility.

1. INTRODUCTION

For several nuclear applications such as neutron spectrum measurements in zero-power experimental nuclear reactors, nuclear material detection, fusion diagnostics etc., there is a need for fast neutron detectors that can cover a wide energy range, especially from a few 10 keV's to 10 MeV, and that can operate in mixed neutron-gamma radiation fields. These neutron spectrometers must have a good energy resolution (<10%), high detection efficiency, be compact and easy to handle, while not being too sensitive to thermal neutrons. Organic scintillators have many of these desirable characteristics, their performance being dependent on the type of scintillating material used: liquid, plastic or crystal. However, organic scintillators are sensitive to gamma rays. Consequently, pulse shape discrimination (PSD) techniques have to be applied to the output signals given by the photomultiplier (PMT) to enable neutron-gamma discrimination. For energies below 1 MeV, the discrimination becomes increasingly difficult due to the weakness of the signals given by the light produced by the scintillators. The organic scintillators commonly used as

references for neutron spectroscopy in mixed fields are liquid scintillators such as the BC501A and NE213 [1], but these devices have limitations in terms of energy performance (poor discrimination below 1 MeV) and operational security (fire hazard). In recent years, organic scintillator technology has been significantly improved and digital acquisition systems are now able to replace the traditional analogic systems. The main advantages are the reduction of the electronic acquisition chain and the possibility to store each event and its information like time and pulse shapes. In this way, it is possible to perform offline studies. This is much more difficult with analog systems, therefore the signal generally needs to be processed as it is generated. Recently, a new organic scintillator based on a Ø25 x 25 mm solution-grown stilbene single crystal was developed [2] and tested at 5 MeV and 565 keV. However, the available measurements were insufficient for establishing the low neutron-energy threshold below which neutron-gamma discrimination becomes impossible. Our goal is therefore to investigate this threshold in mono-energetic neutrons fields between 80 keV and 230 keV.

2. EXPERIMENTAL DETAILS

A. The AMANDE facility

The AMANDE facility of the Institute for radiological protection and nuclear safety (IRSN) at Cadarache is able to produce continuous or pulsed monoenergetic neutrons from ~2 keV up to 20 MeV [3]. The neutron field for this work was produced using the ${}^7\text{Li}(p,n)$ reaction with a proton beam of 2067 keV energy and 1 μA intensity sent on a LiF target of 107 $\mu\text{g}/\text{cm}^2$ thickness. The neutron energy at 0° , i.e. in the proton beam direction is 304 keV, but more neutron energies are actually accessible by varying the detection angle as shown in Table 1. The experimental area (20 x 20 x 16 m^3), is designed so as to limit neutron returns by scattering on the walls.

*Corresponding author:
E-mail: Augusto.DICHICCO@cea.fr
Tel : +33 6 13 18 92 56

Figure 1 Experimental area with the stilbene detector mounted on the mobile arm

Table 1 Angles relative to proton beam direction and corresponding neutron energies

Angle	Neutron energies	Monoenergetic peak FWHM
45°	223 keV	7 keV
60°	179 keV	6 keV
80°	125 keV	5 keV
100°	84 keV	5 keV

Moreover, the presence of a fully-automated transport system can place the detectors at a distance of up to 6 m from the target, and at any angle between 0° and $\pm 150^\circ$. The time-of-flight measurements relies on a pick-up ring placed in the beam at 1 m in front of the target. The ion beam pulsation is performed using a chopper-buncher system [3b] allowing pulse width around 2 ns.

B. Stilbene Organic Scintillator

The solution-grown single crystal of stilbene was produced by InradOptics [4], coupled with a photomultiplier ET 9214B of the ET Enterprise [5]. The whole detector was assembled by SCIONIX [6]. The stilbene is encapsulated in an aluminium canister and is optically coupled to the photomultiplier, which is surrounded by a solid mu-metal shield. When an ionizing particle interacts with the scintillator, light is produced. The signals due to incoming gamma rays are different from those produced by neutrons. Gamma-ray interaction leads to electrons and a low-density charge deposit, which excites atoms principally to singlet states. The relaxation times of excited singlet states are very short (less than 10 ns) and the related light production is called prompt fluorescence. Neutrons, on the other hand, generate protons via elastic scattering on hydrogen, leading to a high-density charge deposit, which excites atoms to single and triplet states. The lifetime of the excited triplet states is longer (more than μs) and the corresponding light emission is called delayed [7]. The output signal is taken from the anode of the photomultiplier and the difference in light

emission is used to discriminate neutrons (signal with longer decay time due to delayed fluorescence) from photons (prompt fluorescence mainly), as illustrated in Figure 2. However, as the signal amplitude is directly related to the energy of the incident particle (neutron or photon), when neutron energy decreases, the baseline variation (see Section C below) has a higher impact on the ratio between prompt and delayed fluorescence, and it becomes increasingly difficult to distinguish between the two types of particles.

Figure 2 Typical shapes of gamma rays and neutrons

C. Digital Acquisition System and Charge Comparison Method

The anode signal from the PMT is directly connected to a DAQ CAEN 5730 [8]. This device has eight channel, 14-bit resolution, 500 MHz sampling rate and a maximum digitizable voltage of 2 V. The online measurements have been performed with the CAEN software CoMPASS [9] using the Digital Pulse Processing – Pulse Shape Discrimination (DPP-PSD) firmware. In order to perform a good online analysis, the DAQ parameters, shown in Figure 3, have to be optimized according to the type of detector and the experimental situation.

Figure 3 Most relevant parameters for the DPP firmware

Once the recorded signal $i(t)$ exceeds a certain threshold value, the firmware uses the Charge Comparison Method (CCM) to calculate the integrals of the pulses over two different time windows, called “short gate” and “long gate”, respectively. These two integral quantities are used in turn to estimate the total light output and the prompt light output. In this way, it is possible to infer a PSD value and discriminate between the detected particles. The equations used to calculate the two integrals are the following:

$$Q_S = \int_{t_0}^{t_s} i(t) dt \quad (1)$$

$$Q_L = \int_{t_0}^{t_L} i(t) dt \quad (2)$$

where t_0 is the start time, called pre-gate, for the integrals calculation, while t_s and t_L correspond respectively to the short and long gates (the optimal values are 34 ns and 414 ns for the stilbene). The baseline is determined by averaging over 256 samples, it is subtracted from the signal during the calculation. The PSD is obtained by the following equation:

$$PSD = \frac{Q_L - Q_S}{Q_L} \quad (3)$$

The two-dimensional (2D) histogram of the PSD for an AmBe source measured with the stilbene is plotted as a function of Q_L (directly related to the energy) in Figure 4 (a). The two areas represent respectively neutrons (high PSD) and gamma rays (low PSD). The PMT was set at -930 V in order to avoid saturation and excessive pile-up counts. The data were handled with ROOT [10].

Figure 4 (a) The PSD vs. Q_L of the AmBe source; (b) Light amplitude for various Q_L when applying a cutoff value to the PSD

Figure 4 (b) shows the total light produced (in black) in log scale. The Compton Edge of the 4.438 MeV gamma ray due to the ${}^9\text{Be}(\alpha, n){}^{12}\text{C}^*$ reaction is clearly visible. By considering a cutoff value of 0.26 in the PSD, it is possible to separate the individual contributions of neutrons ($0.26 < \text{PSD} < 0.6$, in blue) and gamma-rays ($0.06 < \text{PSD} < 0.26$, in red) to the total signal.

D. Time of Flight technique

In the low neutron energy range, the weakness of the signals does not always allow to obtain a clear double-peak PSD profile, and hence setting a cutoff value is difficult. At AMANDE, the time-of-flight technique has been used to help distinguish neutrons from gamma-rays, and to improve the resolution of the PSD. To that aim, the proton beam has been pulsed and the pick-up (PR) ring signal has been connected to the second channel of the DAQ and used as stop signal. The stilbene detector, positioned at a fixed distance $L = 1,907$ m from the target, is connected to the first channel of the digitizer and used as start signal.

Figure 5 Time of flight scheme for neutron-gamma beams

The DAQ, working in coincidence mode, records the start time of each channel. The difference between these start times is determined with a total resolution of less than 3 ns, for gamma-rays and neutrons simultaneously produced in the target by the pulsed proton beam. The neutron time-of-flight between the target and the detector is simply obtained by writing:

$$ToF_{measured} = t_n - t_\gamma + \frac{L}{c} \quad (4)$$

where t_n and t_γ are the neutron and gamma detection times, respectively, as shown in Figure 5.

3. EXPERIMENTAL RESULTS

For the low energy range measurements, the voltage of the PMT was increased up to -1550 V. However, with this fairly-high voltage setting for the PMT, many signals are lost as their amplitude exceeded the 2 V digitizing window, as shown in Figure 6. In addition, there are spurious contributions that give negative values of Q_L and Q_S , or even some where the total light output value is lower than the prompt one. All these contributions had therefore to be removed from the data analysis.

Figure 6 Examples of unusual signals

In Figure 7 (a), the PSD is plotted versus Q_L where it is still possible to discriminate the neutrons from the photons. The upper part of the plot ($PSD > 0.7$) corresponds to signals with bad pulse shape that must be eliminated. In Figure 7 (b), the PSD is plotted versus the time difference between the start and stop from the DAQ, in pulse mode. The areas corresponding to the gamma-rays at 100 ns and to the neutrons at 404 ± 10 ns are clearly visible. Below the neutron structure, it is also possible to see the gamma rays produced

by neutron interactions in the stilbene crystal.

Figure 7 Stilbene data from the DAQ for 230 keV neutrons

In Figure 7 (c), the total pulse height spectra (in blue) is plotted, as well as the derived neutron PHS after selecting the neutron contributions in the time interval $390 < \Delta t < 425$ ns. Figure 8 shows the time distributions recorded in the four neutron fields from Table 1. The monoenergetic peak is clearly visible even at 84 keV, the lowest energy. The corresponding recoil proton energy distributions were obtained by selecting only the events corresponding to the monoenergetic neutrons in the time interval, as shown in Figure 9. These distributions will be used to validate upcoming calculated responses of the stilbene scintillator.

Figure 8 Time of flight distributions for the stilbene

Figure 9 Proton recoil spectra

4. CONCLUSIONS

In this article, the capability of an organic scintillator based on a Ø25 x 25 mm solution-grown stilbene single crystal to measure neutrons with energies below 100 keV has been tested at the AMANDE facility, with the help of the time-of-flight technique to improve the PSD analyses. The results show that, even at such “low” energies, the stilbene detector still has some neutron-gamma discrimination capabilities, which is remarkable. Upcoming work will focus on the study of individual pulse shapes, so as to have more precise criteria to distinguish neutrons and gamma rays. In addition, the crystal anisotropy will be studied in this energy range as it can significantly influence the detector response. The next step of these studies will be to calculate the stilbene detector response matrix with Monte Carlo codes and to compare it with the experimental data.

ACKNOWLEDGMENT

The authors thank the entire AMANDE team of IRSN, especially Dr. Richard Babut, for his collaboration and invaluable assistance during the experiments.

REFERENCES

- [1] H. Klein, “Neutron spectrometry in mixed fields: NE213/BC501A liquid scintillation spectrometers,” *Radiat. Prot. Dosimetry*, vol. 107, no. 1–3, pp. 95–109, Nov. 2003.
- [2] L. Dioni, V. Gressier, G. Nardin, R. Jacqmin, B. Stout, and M. Sumini, “Tests of a solution-grown stilbene scintillator in mono-energetic neutron beams of 565 keV and 5 MeV,” *Nucl. Instrum. Methods Phys. Res.*

Sect. Accel. Spectrometers Detect. Assoc. Equip., vol. 880, pp. 210–215, Feb. 2018.

- [3] V. Gressier *et al.*, “AMANDE: a new facility for monoenergetic neutron fields production between 2 keV and 20 MeV,” *Radiat. Prot. Dosimetry*, vol. 110, no. 1–4, pp. 49–52, Aug. 2004.
- [4] Inforad, “Production of Stilbene for Fast Neutron Detection.”
- [5] “9214B Series - ET-Enterprises Ltd.” [Online]. Available: <http://et-enterprises.com/products/photomultipliers/product/p9214b-series>. [Accessed: 05-May-2019].
- [6] “Homepage Scionix.nl * Scionix,” *Scionix*. [Online]. Available: <https://scionix.nl/>. [Accessed: 05-May-2019].
- [7] *The Theory and Practice of Scintillation Counting*. Elsevier, 1964.
- [8] “DT5730,” *CAEN - Tools for Discovery*. [Online]. Available: <https://www.caen.it/products/dt5730/>. [Accessed: 05-May-2019].
- [9] “CoMPASS,” *CAEN - Tools for Discovery*. [Online]. Available: <https://www.caen.it/products/compass/>. [Accessed: 05-May-2019].
- [10] “ROOT a Data analysis Framework | ROOT a Data analysis Framework.” [Online]. Available: <https://root.cern.ch/>. [Accessed: 06-May-2019].